InControl

Quality – Made in Germany

✓ O$_2$

✓ CO$_2$

✓ °C

Independent measuring device for control and evaluation of CO$_2$, temperature and O$_2$

Features:

- Accurate measuring of CO$_2$, temperature and optional O$_2$

- Single or continuous measurement in arbitrary intervals

- Measurement and documentation of multiple incubators

- Documentation of measurements with date, time and incubator number

- Easy menu navigation

- Data-Download via USB; PC software DataVISUAL`09 incl.

- Continuous operation by Li-Ion battery or mains adapter possible

Scope of supply:

- InControl 1050 including Transport case

- Mains adapter

- Tubing, adapter and condensation trap

- PC software: DataVISUAL'09 (on USB-Stick)

- USB data cable

REF 14709

Adress:
Labotect GmbH
Labor-Technik-Göttingen
Kampweg 12
D-37124 Rosdorf

Postal adress:
Labotect GmbH
Labor-Technik-Göttingen
Postfach 200212
D-37087 Göttingen

Phone   +49 551 50 50 10
Fax   +49 551 50 50 111
eMail: info@labotect.com
Internet: www.labotect.com

Court of register: Göttingen: HRB 1394
WEEE-Reg-No: DE 86039820
Managing director:
Delia Schinkel-Fleitmann
Dr. Friederike Fleitmann

Technical data

Case:
- Plastic (ABS)

CO_2 measurement:
- Dual beam IR sensor
- Range: 0 - 10% CO_2
- Display definition: 0.1% CO_2
- Accuracy:
  - 0 - 6% CO_2 ± 0.2% CO_2
  - 6 - 10% CO_2 ± 0.3% CO_2
- Sample volume per measurement (2 min.): approx. 0.8 l

Temperatur measurement:
- Platinum-temperature sensor PT1000
- Range: 0 - 100°C
- Display definition: 0.1°C
- Accuracy:
  - 20 - 50°C ± 0.2°C
  - 50 - 100°C ± 0.3°C

O_2 measurement:
- Galvanic sensor
- Range: 0 - 100% O_2
- Display definition: 0.1% O_2
- O_2-sensor not included

Data logger:
- Recording of measurement logging in arbitrary intervals
- Output of measurements via:
  - backlit LCD
  - PC software Data Visual'09 (via USB data cable)
- Interval for data logging: 15 - 120 min. adjustable in steps of 5 minutes
- storage of max. 1008 measurements with date and time

Power supply:
- Rechargeable Li-Ion battery
- Power supply 100 – 240 V AC, 5 V DC 1.5 A

Operation time:
- Use of Rechargeable battery approx. 4 hrs
- Time to recharge battery approx. 4 hrs

Optional accessories:

Sensor for O_2 measurement

REF   10556

Flat tape sensor PT1000 for temperature measurement inside Labo C-Top or other benchtop incubators

REF   15064

Surface sensor for temperature measurement

REF   10537

TÜV SÜD
EN ISO 13485
certified company

CE

Rev. 4_11/2013